

EC Design Examination Certificate

Certificate No.:
12475-2018-CE-IND-NA-PS

Project No.:
PRJC-03148-2007-PRC-IND

Valid Until:
25 April 2023

This is to certify that:
Sterile Disposable Medical Devices

Manufactured by:

BL Lifesciences Pvt. Ltd.
28-D, Sector 31, Ecotech I, Greater Noida,
Gautam Budh Nagar, U.P., India

Has been assessed with respect to:

**Examination of the design of the product as described in Annex II
section 4 (Module B1) of Council Directive 93/42/EEC on Medical
Devices, as amended**

and found to comply.

Further details of the product(s) and conditions for certification are given overleaf.

Place and Date:
Høvik, 27 April 2018



For:
DNV GL NEMKO PRESAFE AS

Tone Kolpus

The Certificate has been digitally signed.
See www.presafe.com/digital_signatures for more info

Notice: The Certificate is subject to terms and conditions as set out in the Certification Agreement. Failure to comply may render this Certificate invalid.

EC Design Examination Certificate

Certificate No.:
12475-2018-CE-IND-NA-PS

Project No.:
PRJC-03148-2007-PRC-IND

Valid Until:
25 April 2023

Jurisdiction

Application of Council Directive 93/42/EEC of 14 June 1993, adopted as “Forskrift om Medisinsk Utstyr” by the Norwegian Ministry of Health and Care Services.

Certificate history:

Revision	Description	Issue Date
	Original Certificate	2018-04-25

Products covered by this Certificate:

Type of medical device and identification no.: Sterile Disposable Medical Devices	Medical Device Class: III	GMDN code:
Short description of the Medical Device: CVC is used for catheterization of Superior Vena cava using Seldinger technique in infusion therapy or parenteral nutrition, application of vein irritation solutions, for intermittent or continuous monitoring of central venous pressure & for blood sampling. Method of Sterilization: Ethylene Oxide		

Terms and conditions

The certificate is subject to the following terms and conditions:

- Any producer (see 2001/95/EC for a precise definition) is liable for damage caused by a defect in his product(s), in accordance with directive 85/374/EEC, as amended, concerning liability of defective products.
- The certificate is only valid for the products and/or manufacturing premises listed above.
- The Manufacturer shall inform Presafe of any intended change of the products detailed above and Presafe will assess the changes and decide if the certificate remains valid.

The following may render this Certificate invalid:

- Changes in the design of the products to which this Certificate refers.
- Changes in requirements of the scheme to which this Certificate refers.

Conformity declaration and marking of product

This Certificate must be accompanied with a valid EC Certificate Full Quality Assurance System.

When meeting with the terms and conditions above, the producer may draw up an EC declaration of conformity and legally affix the CE mark followed by the Notified Body identification number of Presafe.

End of Certificate